

#### PRODUCT SPECIFICATION

# Patterson® TactileGuard Ultra 3G™ Nitrile Examination Gloves

### **PRODUCT**

Nitrile examination glove Medical grade Non-sterile Powder-free Textured fingertips

# **COUNTRY OF ORIGIN**

Malaysia

# **INTENDED USE**

This is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.

### **MATERIAL**

Synthetic Nitrile Rubber (NBR)

# **OUTER SURFACE**

No donning powder used

# **COMPONENTS**

Synthetic rubber nitrile (NBR)
Titanium Oxide
Sulfur
Organic accelerators (carbamates, thiazolates)

#### SHAPE

Straight fingers
Thumb and fingers in one plane
Ambidextrous

### **CUFF**

Beaded (rolled rim)

### **COLOR**

Blue

#### SIZES

Extra small (XS), small (S), medium (M), large (L), extra large (XL)

# **MARKING**

Packaging marked to designated size (gloves not marked)

### **PACKAGING AND LABELING**

Reorder Number 088-5714, 088-5723, 088-5731, 088-5749, 088-5756 300 pieces per box (250 pieces XL), 3000 pieces per case (2500 pieces XL)

### **CONTROL NUMBER (LOT NO.)**

Each packing unit (dispenser box) and outer carton bears a control number

EXAMPLE: 130913180115

Key: 13 ..... Production Year
09 ..... Production Month
13 ..... Production Day
180115 ..... Batch Number

#### QUALITY CHARACTERISTICS\*

All listed standards are used in their latest edition

| DESCRIPTION                           | SPECIFICATION                 | ASSURANCE ACTION |
|---------------------------------------|-------------------------------|------------------|
| <u>Dimensions</u>                     |                               | ASTM D 6319-00a  |
| Overall length                        | 220 mm min (S)                |                  |
|                                       | 230 mm min (S, M, L, XL)      |                  |
| Width                                 | 70 mm +/- 10 mm (XS)          | ASTM D6319       |
|                                       | 80 mm +/- 10 mm (S)           |                  |
|                                       | 95 mm +/- 10 mm (M)           |                  |
|                                       | 110 mm +/- 10 mm (L)          |                  |
|                                       | 120 mm +/- 10 mm (XL)         |                  |
| Thickness (single wall)               | Finger: 0.05 mm/2.0 mils min. |                  |
|                                       | Palm: 0.05 mm/2.0 mils min.   |                  |
| Physical properties                   |                               | ASTM D 412       |
| Tensile strength (before/after aging) | 14 MPa min.                   |                  |
| Elongation (before aging)             | 500% min.                     | ASTM D 412       |
| (after aging)                         | 400% min.                     |                  |

<sup>\*</sup>minimum acceptable requirements of ASTM and FDA where applicable. Gloves offered by Sempermed USA, Inc. meet or exceed these physical requirements. Independent laboratory test results are available upon request.

### PERFORMANCE REQUIREMENTS FOR QUALITY CHARACTERISTICS

For reference purpose in accordance with ISO 2859 "Sampling Procedures for Inspection by Attributes". All standards listed in this specification are applied to medical gloves non-sterile.

### INTERNAL ATTRIBUTIVE RELEASE INSPECTION

Sampling for examination in accordance with ISO 2859 Unit for *inspection*: one (1) glove If several defects are found on one glove, only the most serious defect (i.e. lowest category) is evaluated. The acceptance criteria is based on the number of defectives observed in a sample.

# **FINAL GLOVE RELEASE**

Assurance action

ASTM D 6319 "Standard Specification for Nitrlie Examination Gloves for Medical Application" ASTM D 5151: "Standard Test Method for Detection of Holes in Medical Gloves"

# Sampling inspection and final release information

*Major defects:* highest concern non-conformities which prevent correct use of the product. AQL 1.5 (inspection level GI for leaks)

*Minor defects*: non-conformities of a lesser degree of concern, which do not prevent correct use of the product. AQL 4.0 (inspection level GI for visual defects aggregated)

### PACKAGING, MARKING, GOOD DELIVERY INSPECTION

#### **Assurance Action**

Set-up and patrol inspection at packaging Supervision of vehicle or vessel loading

Sempermed USA, Inc. is a certified participant in C-TPAT (U.S. Customs-Trade Partnership Against Terrorism)

#### **GOOD MANUFACTURING PRACTICE**

The gloves are manufactured in compliance with ISO 9001, ISO 13485, and US FDA 21 CFR part 820

#### MICROBIOLOGICAL CLEANLINESS CONTROL

The bioburdens of the finished gloves are monitored and recorded. Unusual contaminants are identified.

**CAUTION:** Non-sterile examination gloves are used in a variety of circumstances, including procedures where the surface of the glove contacts wounds, body cavities, or other possible routes of contamination. If conditions warrant, the user may wish to minimize the risk of infection. In this case we recommend the decontamination of the gloves prior to use by disinfectants or other effective methods.

#### **STORAGE**

According to ISO 2230 for Vulcanized Rubber Store in a dry, ventilated area Avoid direct sunlight, fluorescent lighting, heat, and moisture Do not store above 100° F (38° C) as this will lead to accelerated aging Long-term storage can result in pleats and stickiness

**END OF DOCUMENT** 

